

**MP Biomedicals, LLC**

29525 Fountain Parkway  
Solon, Ohio 44139

Telephone: 440/337-1200  
Toll Free: 800/854-0530

Fax: 440/337-1180  
web: [www.mpbio.com](http://www.mpbio.com)

## Certificate of Analysis

**Product Description:** 0.25% Trypsin (1:250) and 1mM EDTA in Modified HBSS w/o Ca/Mg

**Catalog Number:** 1689649; 1689654

**Lot:** 3877B

**Formula:** N/A

**CAS #:** N/A

**Physical Description:** Reddish-orange Solution

**Formula Weight:** N/A

**Storage:** -20°C

Test	Specification	Result
pH	7.0 - 7.3	7.1
Osmolarity	270 - 305 mOsm	282 mOsm
Functionality Test	Pass	Pass
Sterility Test	Pass	Pass
Mycoplasma	Not detected	Pass

**EXPIRATION DATE:** 3/31/00

Approved - 09/16/99 by Joe Dietz

MP Biomedicals, LLC.

Technical Director

This is an electronically generated document

<mailto:biotech@mpbio.com>

<http://www.mpbio.com>

Online Ordering, MSDSs, certificates of analysis and data sheets now available on our web site

Technical Service: 1-800-279-5490 (440-337-1200) Customer Service: 1-800-854-0530 (440-337-1200)