

 29525 Fountain Parkway
 Telephone: 440/337-1200
 Fax: 440/337-1180

 Solon, Ohio 44139
 Toll Free: 800/854-0530
 web: www.mpbio.com

Certificate of Analysis

Product Description: Mono-poly Resolving Medium_

Catalog Number: 1698049_

Lot: 16980004_

Formula: Mixture of ficoll and hypaque Formula Weight: n/a

CAS #: n/a Storage: Room temperature

Physical Description: Clear, colorless liquid

Test	Specification	Results
Osmolality	380 - 471 mOsm/kg	405 mOsm
Specific Gravity	1.112 - 1.116 g/ml	1.113 g/ml
Sterility	Passes	Passes
Mycoplasma	Tested Negative	Negative
рН	For Information Only	7.1
% of Mononuclear Cells	≥ 90%	95.3%
% of Polymorphonuclear Cells	≥ 89%	96.7%
Expiration Date	m/day/yr	02/28/00

Approved - 03/02/99 by Joe Div

Technical Director