

**MP Biomedicals, LLC**

29525 Fountain Parkway  
Solon, Ohio 44139

Telephone: 440/337-1200  
Toll Free: 800/854-0530

Fax: 440/337-1180  
web: [www.mpbio.com](http://www.mpbio.com)

## Certificate of Analysis

**Product Description:** FastCellular Senescence Detection Kit

**Catalog Number:** 2690301

**Lot:** S5177

**Formula:** N/A

**CAS #:** N/A

**Physical Description:** Kit

**Formula Weight:** N/A

**Storage:** protect from light, 2 - 8°C

Test	Specification	Result
Identity Test	Passes	Passes
Purity (SPiDER-bGal; HPLC)	Report Result	96.7%
Purity (Bafilomycin A1; HPLC)	Report Result	97.8%
Fluorescence Spectrum	Passes	Passes

Recommended Retest Date (MM/DD/YYYY): 03/27/2022

12/03/2020 - John Huang, PhD

MP Biomedicals, LLC.

Quality Control Manager

This is an electronically generated document

<mailto:biotech@mpbio.com>

<http://www.mpbio.com>

Online Ordering, MSDSs, certificates of analysis and data sheets now available on our web site

Technical Service: 1-800-279-5490 (440-337-1200) Customer Service: 1-800-854-0530 (440-337-1200)