

**MP Biomedicals, LLC**

29525 Fountain Parkway  
Solon, Ohio 44139

Telephone: 440/337-1200  
Toll Free: 800/854-0530

Fax: 440/337-1180  
web: [www.mpbio.com](http://www.mpbio.com)

## Certificate of Analysis

**Product Description:** NP-40 Lysis Buffer with Glycerol (2X)

**Catalog Number:** 823211

**Lot:** S7216

**Formula:** N/A

**CAS #:** N/A

**Physical Description:** Clear, Colorless Liquid

**Formula Weight:** N/A

**Storage:** 2 - 8°C

| Test          | Specification | Result |
|---------------|---------------|--------|
| Identity Test | Passes        | Passes |

Tris Base/Tris-HCl: 100mM

Sodium Chloride: 300mM

EDTA Disodium Dihydrate: 10mM

IGEPAL CA-630: 2%

Glycerol: 10%

pH(1x)=7.4 ± 0.15

Recommended Retest Date (MM/DD/YYYY): 10/05/2022

10/06/2021 - John Huang, PhD

MP Biomedicals, LLC.

Quality Control Manager

This is an electronically generated document

<mailto:biotech@mpbio.com>

<http://www.mpbio.com>

Online Ordering, MSDSs, certificates of analysis and data sheets now available on our web site

Technical Service: 1-800-279-5490 (440-337-1200) Customer Service: 1-800-854-0530 (440-337-1200)